



PSB Singapore

TEST CERTIFICATE

This Certificate is issued to

Tan Soon Huah Gas Supply Pte Ltd
 15 Defu Lane 11
 Singapore 539171

FOR

Product: Diaphragm Gas meter

Brand/Model: DAESUNG / G4R

Detail: Maximum flow rate (Q_{max}): 6.0 m³/h
 Minimum flow rate (Q_{min}): 0.040 m³/h
 Maximum working pressure (P_{max}): 1.5 bar
 Nominal value of the cyclic volume (V): 1.7 dm³


Specification: OIML R 137-1:2006 – Gas meters – Part 1: Requirements

Test Report: 7191016766-MEC12/LGH

Date of Test Report : 28 Mar 2012

Summary

A sample of product submitted was tested and was found to comply with the test requirements of the above standard(s).


 Vice President (MEC)
 TÜV SÜD PSB Pte Ltd

Certificate No:	Date of Original Issue:	Date of Last Revision:	Date of Expiry:
02840	05/04/2012		04/04/2014

This Certificate is part of a full report and should be read in conjunction with it. This Certificate remains the property of TÜV SÜD PSB Pte Ltd and shall be returned upon request. The use of this Certificate is subjected to the terms and conditions of the Test Certification Scheme.

Note: This Certificate is issued pursuant to the terms set out overleaf.